



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

ENV Services, Inc./Pro-Lab
4758 Research Drive, San Antonio, TX 78240

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Mechanical Testing
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Initial Accreditation Date:

August 30, 2016

Issue Date:

January 18, 2023

Expiration Date:

February 28, 2025

Accreditation No.:

78961

Certificate No.:

L23-50-2

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjilabs.com



Certificate of Accreditation: Supplement

ENV Services, Inc./Pro-Lab

4758 Research Drive, San Antonio, TX 78240
 Contact Name: Michael Parente Phone: 215-997-5080

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED
Mechanical ^o	Fume Hoods, Bio-Safety Cabinets, Glove Boxes, Clean Room, Etc.	Chemical Fume Hood	ASHRAE 110, SEFA 1, ANSI/AIHA Z.95, ENV Method WI-4380 & WI-4381
		Ductless Fume Hood	SETA 1, IEST-RP-CC034.3
		Class I BioSafety Cabinet	WI-4389
		Class II, Type A1 BioSafety Cabinet	NSF/ANSI 49 (Except Vibration), WI-4374
		Class II, Type A2 BioSafety Cabinet	NSF/ANSI 49 (Except Vibration)
		Class II, Type B1 BioSafety Cabinet	NSF/ANSI 49 (Except Vibration), WI-4374
		Class II, Type B2 BioSafety Cabinet	NSF/ANSI 49 (Except Vibration), WI-4374
		Class III BioSafety Cabinet	IEST-RP-CC028/.1, ISO 14644-7, WI-4382
		Glove Box	IEST-RP-CC028.1, ISO 14644-7, AGS-G001
		Laminar Flow Glove Box	IEST-RP-CC028.1, ISO 14644-7, AGS-G001
		Barrier Isolator	IEST-RP-CC028.1, ISO 14644-7, AGS-G001
		Laminar Flow Hood	IEST-RP-CC002.3, WI-4387
		Laminar Flow Module	IEST-RP-CC002.3, WI-4387
		Terminal HEPA Filter	IEST-RP-CC002.3, IEST-RP-CC034.2, WI-4387
		Clean Room	ISO 14644, Part 1 & 2, IEST-RP-CC006.3, FDA Guidance for Industry 2004, WI-1146
		Certificate of Sterile Compounding Facilities and Aseptic Isolators	USP797, CETA CAG-001, CETA CAG-002, CETA CAG-003
		Ovens, Furnaces, Freezers, Environmental Chambers-Temperature Uniformity Test ^o	-195 °C to 1372 °C 20 % RH to 90 % RH

- The presence of a superscript O means that the laboratory performs testing of the indicated parameter onsite at customer locations. Example: Outside Micrometer^o would mean that the laboratory performs this testing onsite at the customer's location.